

Crowcon Technical Note

Document Reference: GEN092

Document applies to: T3 and Gasman

Release Date: 08/07/2019

Subject: Servicing and replacement of H2S sensors in T3 and Gasman



We are changing our H2S sensor to a new version in our T3 and Gasman portable gas detector models.

Please use the below new part code to order replacement H2S sensors for T3 in ATEX regions only and for Gasman in ATEX and INMETRO regions only

New: S011812/S

- This applies to existing T3 & Gasman units already in the field as well as new orders
- Calibration procedure and gas requirements and procedures remain the same as before
- Pricing remains the same as the existing version

****Please note for INMETRO and CSA certified T3 H2S sensors and CSA certified Gasman H2S sensors you must continue to use the existing part code****

Existing: S011421/M

For more information please contact +44 (0)1235 557711, technicalsupport@crowcon.com

